

G.A.247447

Collaborative Project of the 7th Framework Programme



ain



D3.2: Sensors, Signals and Signal Processing, SoA

Guger Technologies OG
v1.2_PU

30/06/2010

www.BrainAble.org

The financial support of the European Commission is gratefully acknowledged. Material in this document reflects only the author's view and the Commission is not liable for any use that may be made of the information contained herein.

Its owner is not liable for damages resulting from the use of erroneous or incomplete confidential information.

Document Information

Project Number	247447	Acronym	BrainAble
Full title	Autonomy and social inclusion through mixed reality Brain-Computer Interfaces: connecting the disabled to their physical and social world		
Project URL	http://www.BrainAble.org		
EU Project officer	Jan Komarek		

Deliverable	Number	3.2	Title	Signal, sensors and signal processing (SoA)
Work package	Number	3	Title	Multimodal sensing and monitoring system for BNCI, affective and biometric signals

Date of delivery	Contractual	PM06	Actual	PM06
Status	Reviewed		final <input checked="" type="checkbox"/>	
Nature	Prototype <input type="checkbox"/> Report <input checked="" type="checkbox"/> Dissemination <input type="checkbox"/> Other <input type="checkbox"/>			
Dissemination Level	Public <input checked="" type="checkbox"/> Consortium <input type="checkbox"/>			

Authors (Partner)	g.tec, TUG, UPF			
Responsible Author	Rupert Ortner	Email	gruenbacher@gtec.at	
	Christoph Guger Engelbert Grünbacher			
	Partner	g.tec	Phone	+43 7251 22240 16

Abstract (for dissemination)	This report presents a state of the art of sensors, signals and signal processing for measuring EEG and non – EEG data. Due to the wide range of biosignal analysis this document is focused to sensors and signals that are useful within BrainAble.
Keywords	Biosignals, Sensors, BCI, Biosignal amplifier, Signal processing

Version Log			
Issue Date	Version	Author	Change
28.02.2010	0.0	TUG, g.tec, UPF	Table of physiological sensors as a basis for the document
27.4.2010	0.1	Guger, Ortner, Grünbacher	Initial Document
11.6.2010	1.0	Guger, Ortner, Grünbacher	First Draft for review
15.6.2010	1.1	Aleksander Valjamae (UPF)	Video tracking system
29.6.2010	1.2	“ + PMO (formatting)	Final version released to the P.O.
30.6.2010	v1.2_PU	PMO	Publishable version: suppression of project' outlines

The information in this document is provided as is and no guarantee or warranty is given that the information is fit for any particular purpose. The user thereof uses the information at its sole risk and liability. Its owner is not liable for damages resulting from the use of erroneous or incomplete confidential information.

Index

DOCUMENT INFORMATION	2
INDEX.....	3
LIST OF FIGURES	4
LIST OF TABLES	4
1 INTRODUCTION	5
GENERAL NOTE.....	6
2 SENSORS FOR BIOSIGNAL MEASUREMENTS	6
2.1 SENSORS AND CAPS FOR MEASURING EEG SIGNALS.....	6
2.1.1 <i>Electrodes</i>	6
2.1.2 <i>Electrode Caps</i>	7
2.2 PULSE RATE SENSORS	7
2.3 DISPOSABLE ELECTRODE (ECG, EMG)	8
2.4 FINGER ELECTRODE (ELECTRODERMAL ACTIVITY).....	8
2.5 RESPIRATION BELT / NOISE FLOW SENSOR.....	8
3 BIOSIGNAL AMPLIFIER CONCEPT	9
3.1 HARDWARE DESCRIPTION	9
3.2 PROGRAMMING ENVIRONMENT.....	13
4 SIGNALS AND SIGNAL PROCESSING ALGORITHMS	14
4.1 ECG.....	14
4.1.1 <i>Heart Rate (HR)</i>	14
4.1.2 <i>Heart Rate Variability (HRV)</i>	15
4.2 EEG.....	16
4.2.1 <i>Slow Cortical Potentials</i>	17
4.2.2 <i>P300</i>	17
4.2.3 <i>SSVEP</i>	18
4.2.4 <i>Motor imagery</i>	18
4.3 EOG AND EYE MOVEMENTS	19
4.4 EMG	19
4.5 ELECTRODERMAL ACTIVITY (EDA)	19
4.6 RESPIRATION.....	20
4.7 TRACKING BODY MOVEMENTS OF USERS	20
4.7.1 <i>Video based tracking in VR applications (initiated by UPF)</i>	20
4.7.2 <i>Hand-held devices for gesture tracking</i>	21
5 REFERENCES	23
6 LIST OF KEY WORDS/ABBREVIATIONS	29

List of figures

FIGURE 1: ELECTRODES FOR EEG MEASUREMENTS.....	7
FIGURE 2: ELECTRODE CAPS ACCORDING THE 10/20 ELECTRODE SYSTEM	7
FIGURE 3: PULSE RATE SENSOR WITH BIOSIGNAL AMPLIFIER	8
FIGURE 4: DISPOSABLE ELECTRODES WITH ACTIVE CLIPS.....	8
FIGURE 5: DISPOSABLE ELECTRODES.....	8
FIGURE 6: PASSIVE CLIPS FOR DISPOSABLE ELECTRODES.....	8
FIGURE 7: GALVANIC SKIN RESPONSE SENSOR FOR FINGER MOUNTING (FINGER ELECTRODES)	8
FIGURE 8: RESPIRATION AND FLOW SENSOR WITH BIOSIGNAL AMPLIFIER AND USER.....	9
FIGURE 9: BIOSIGNAL AMPLIFIERS	9
FIGURE 10: BLOCK DIAGRAM OF THE BIOSIGNAL AMPLIFIER G.USBAMP.	10
FIGURE 11: SCALP ELECTRODE MONTAGE WITH BIPOLAR (A) AND MONOPOLAR RECORDINGS (B) ...	12
FIGURE 12: ECG ELECTRODE WITH 64 ELECTRODES.....	13
FIGURE 13: TYPICAL EEG TRACES FOR OSCILLATORY COMPONENTS.....	17

List of tables

TABLE 1: SENSOR TECHNIQUES FOR BIOSIGNAL MEASUREMENTS (BY G.TEC, TUG AND UPF)	6
TABLE 2: TECHNICAL KEY PROPERTIES OF BIOSIGNAL AMPLIFIERS FOR BCI OPERATION.....	11
TABLE 3: FREQUENCY RANGE OF THE DIFFERENT EEG BANDS	16



1 Introduction

This document presents a state of the art of sensors, signals and signal processing for measuring EEG and non – EEG data. Due to the wide range of biosignal analysis we limit ourselves to sensors and signals that are useful within BrainAble.

BrainAble will research, conceive, design, implement and validate an ICT-based human-computer interaction (HCI) composed of brain/neuronal-computer interaction (BNCI) sensors combined with affective computing and virtual environments to overcome two main shortcomings of people with disabilities related to:

- inner functional independence for daily life activities: (HCI connected to accessible and interoperable home and urban automation)
- outer social inclusion: HCI connected to advanced and adapted social network services.

In order to find out the best pathways of the BNCI system this document presents an overview about existing signals and signal processing algorithm which may be useful for the HCI system to be developed. Therefore the following fields have turned out to be of interest within BrainAble and therefore will be considered in this state of the art. The main important useful sensors can be subdivided into sensors for measuring

- Electroenzephalogram (EEG)
- Electrocardiogram (ECG)
- Respiration
- Electromyogram (EMG)
- Electrooculogram (EOG)

It should be mentioned that this document gives an overview of all categories. If the reader is more focused on BCI techniques (hence EEG), then the reader should be referred to the public deliverable D3.1 which presents a state of the art of BCIs.

The first part of the document includes an overview of sensors and sensor types that are applied for measurements in the above mentioned categories. It is separated into sensors for EEG measurements and sensors for non-EEG measurements. In BrainAble a BNCI and not only a BCI system should be developed. Therefore it is crucial to consider non-EEG data as well.

The next part contains an overview about hardware issues, about biosignal amplifiers and about their use. This section will explain the state of the art of hardware and how this hardware can be exploited to be applied in a research environment. Therefore the biosignals have to be easily measurable and processable. Different system architectures present the actual possibilities of implementing prototyping environments.

The final part of the document contains an overview about the literature in the field of signal processing algorithms that are applied for analysing *e.g.* the ECG or the EEG. The main problem of signal processing is how to extract the somehow included information from the sensor signals.

1.1 General note

Please note that publication and distribution of the whole document or of parts of the document need a written agreement.

2 Sensors for biosignal measurements

State of the art Sensors for biosignal measurements, which may be applied for BrainAble are summarized in Table 1. In general we can subdivide in sensors for measuring EEG data and non-EEG data.

Table 1: Sensor techniques for biosignal measurements (by g.tec, TUG and UPF)

Biosignals	Sensor Type	Sensor Location	Min sampling frequency Hz	Optimal sampling frequency Hz
EEG signals	Sintered AG/AgCl electrodes	Scalp (along the international 10/20 electrodes system)	256	256
Pulse (Heart rate)	Photoelectric sensor	Finger, earlobe	64	256
ECG (Heart rate)	Disposable electrode	<i>e.g.</i> Einthoven I/II, Goldberger, Wilson recording	256	1024
Electrodermal activity	Finger electrode	Hand, foot, forehead	32	32
Respiration	Belt / Nose flow sensor	Thorax, abdominal	32	32
Facial EMG	Disposable electrode	Face	256	2048
EMG	Disposable electrode	Hand, leg	256	2048
EOG	Sintered AG/AgCl electrodes	Vertical/horizontal/diagonal eye	128	256

In the following we describe actual solutions mainly provided by g.tec for measuring the signals mentioned in Table 1.

2.1 Sensors and caps for measuring EEG signals

2.1.1 Electrodes

For EEG measurements normally single disk electrodes made of gold or Ag/AgCl are used (see Figure 2). Gold electrodes are maintenance free and have a good frequency response for EEG, EMG or ECG measurements. For DC derivations with EEG frequencies below 0.1 Hz Ag/AgCl electrodes perform better than gold electrodes. Passive electrodes consist only of the disk material and are connected with the electrode cable and a 1.5 mm medical connector to the biosignal amplifier. Active electrodes have a pre-amplifier with gain 1-10 inside the electrode which makes the electrode less sensitive against environmental noise such as power line interference and cable movements. Because of this fact, active electrodes also work if the electrode-skin impedance is higher than for passive electrodes (should be below 10 kOhm). Active

electrodes have system connectors to supply the electronic components with power. Figure 1A, Figure 1B and Figure 1C show EEG electrodes that can be fitted into EEG caps, Figure 1D shows an ECG/EMG electrode which is placed close to the muscle/heart. Electrodes of type A and D can also be used for EOG recordings.




Figure 1: Electrodes for EEG measurements

A: Active single electrode with multi-pole connector; B: active gold electrode with multi-pole connector; C: screw-able passive gold electrode; D: active ECG electrode with disposable Ag/AgCl electrode

2.1.2 Electrode Caps

EEG electrodes are normally distributed on the scalp according to the international 10-20 electrode system (see Ref. 17). Therefore, the distance from the Inion to the Nasion is first measured. Then, electrode Cz on the vertex of the cap is shifted exactly to 50 % of this distance, as indicated in Figure 2A. Figure 2B shows a cap with 64 positions. The cap uses screwable single electrodes to adjust the depth and optimize electrode impedance. Each electrode has a 1.5 mm safety connector which can be directly connected to the biosignal amplifiers. Active electrodes have system connectors to supply the electronic components with power. There are two main advantages of a single electrode system: (i) if one electrode breaks down it can be removed immediately and (ii) every electrode montage can be realized easily. The disadvantage is that all electrodes must be connected separately each time. Hence, caps are also available with integrated electrodes. All the electrodes are combined in one ribbon cable that can be directly connected to system connectors of the amplifiers. The main disadvantage is the inflexibility of the montage, and the whole cap must be removed if one electrode breaks down.




Figure 2: Electrode caps according the 10/20 electrode system

A: Electrode positioning according to the 10/20 electrode system. B: Electrode cap with screwable single passive or active electrodes. C: Electrode cap with build-in electrodes with a specific montage. D: Electrode cap with active electrodes.

2.2 Pulse rate sensors

A photoelectric sensor can be used to monitor the pulse waves of subjects. The best places to fix it are on the finger or the ear. By detection of the single waves also the heart rate can be extracted from the signal.



Figure 3: Pulse rate sensor with biosignal amplifier

2.3 Disposable electrode (ECG, EMG)

Disposable Silver/Silver-chloride electrodes are convenient for measuring ECG, EMG and facial EMG. They are pre-gelled and self-adhesive, hence they can be fixed onto the subjects skins within seconds, ready to use. Similar to EEG electrodes, it is possible to use active clip connectors, to pre-amplify the signal.



Figure 4: Disposable electrodes with active clips



Figure 5: Disposable electrodes



Figure 6: Passive clips for disposable electrodes

2.4 Finger electrode (electrodermal activity)

This sensor measures the conductance of the skin. Therefore sintered electrodes are fixed with Velcro straps on 2 fingers. A small current is applied to the skin across the electrodes and the resulting voltage drop is measured. The varying output signal of the sensor is proportional to changes in skin conductance.



Figure 7: Galvanic skin response sensor for finger mounting (finger electrodes)

2.5 Respiration belt / Noise flow sensor

For measuring the respiration two types of sensors are common. For the first one an elastic belt has to be fixed around the subject's chest. Inside the belt are piezo elements who convert the expansion (stress) of the belt into an electrical signal. Hence, during expiration, the chests volume will decrease and also the expansion of the belt. Needless to say that it behaves the other way during inspiration.

The second sensor for monitoring respiration is a flow sensor, fixed near the nose and mouth. It is designed to measure the change of temperature during inspiration versus expiration. This is usually done by using a thermocouple.




Figure 8: Respiration and flow sensor with biosignal amplifier and user

3 Biosignal amplifier concept

3.1 Hardware description

One of the key components of a physiological recording and analysis system is the biosignal amplifier. Figure 9 illustrates 3 different devices with different specific key features. g.BSamp is a stand-alone analog amplifier which amplifies the input signals to ± 10 V. The output of the amplifier is connected to a data acquisition board (DAQ) for analog to digital conversion (ADC). g.MOBllab+ is a portable amplifier that transmits already digitized EEG data via Bluetooth to the processing unit. g.USBamp sends the digitized EEG via USB to the processing unit.



Figure 9: Biosignal amplifiers

A: 16 channel stand-alone amplifier g.BSamp. B: 8 channel wireless amplifier g.MOBllab+. C: 16 channel amplifier g.USBamp with USB connection.

A block diagram of g.USBamp is given in Figure 10. This device has 16 input channels, which are connected over software controllable switches to the internal amplifier stages and anti-aliasing filters before the signals are digitized with sixteen 24 Bit ADCs. The device is also equipped with digital to analog converters (DAC) enabling the generation of different signals like sinusoidal waves, which can be sent to the inputs of the amplifiers for system testing and calibration. Additionally, the impedance of each electrode can be checked by applying a small current via the individual electrodes and measuring the voltage drops. All these components are part of the so-called applied part of the device, as a subject or patient is in contact to this part of the device via

the electrodes. All following parts of the device are separated via optical links from the subject/patient.

The digitized signals are passed to a digital signal processor (DSP) for further processing. The DSP performs an over-sampling of the biosignal data, band pass filtering, Notch filtering to suppress the power line interference and calculates bipolar derivations. These processing stages eliminate unwanted noise from the signal, which helps to ensure accurate and reliable classification. Then the pre-processed data is sent to a controller which transmits the data via USB 2.0 to the PC. One important feature of the amplifier is the over-sampling capability. Each ADC is sampling the data at 2.4 MHz. Then the samples are averaged to the desired sampling frequency of e.g. 128 Hz. Here a total of 19.200 samples are averaged, which improves the signal to noise ratio by the square root of 19.200 = 138,6 times.




Figure 10: Block diagram of the biosignal amplifier g.USBamp.
The applied part is surrounded by the dark gray frame.

For EEG or ECoG recordings with many channels, multiple devices can be daisy chained. One common synchronization signal is utilized for all ADCs, yielding a perfect non delayed acquisition of all connected amplifiers. This is especially important for evoked potential recordings or recordings with many EEG channels. If only one ADC with a specific conversion time is used for many channels, then a time lag between the first channel and the last channel could be the result (e.g. 100 channels * 10 μ s = 1 ms). Important is also that biosignal acquisition systems provide trigger inputs and outputs to log external events in synchrony to the data or to send trigger information to other external devices such as a visual flash lamp. Digital outputs can also be used to control external devices such as a prosthetic hand or a wheelchair. An advantage here is to scan the digital inputs together with the biosignals to avoid time-shifts between events and physiological data. A medical power supply that works with 220 and 110 V is required for physiological recording systems that are used mainly in the lab. For mobile applications like the controlling a wheelchair, amplifiers which run on battery power are also useful.

Table 2 compares key technical properties of the 3 amplifiers shown in Figure 3 (g.BSamp, g.MOBilab+ and g.USBamp). The most important factor in selecting an appropriate amplifier is whether non-invasive or invasive data should be processed. For invasive recordings, only devices with an applied part of type CF are allowed. For EEG measurements, both BF and CF type devices can be used. The difference here is the maximum allowed leakage current. Leakage current refers to electric current that is lost from the hardware, and could be dangerous for people or equipment. For both systems, the character F indicates that the applied part is isolated from the other parts of the amplifier. This isolation is typically done based on opto-couplers or isolation amplifiers. For a BF device, the ground leakage current and the patient leakage current must be $\leq 100 \mu\text{A}$ according to the medical device requirements, such as IEC 60601 or EN 60601. These refer to widely recognized standards that specify details of how much leakage current is allowed, among other details. For a CF device, the rules are more stringent. The ground leakage current can also be $\leq 100 \mu\text{A}$, but the patient leakage current must be $\leq 10 \mu\text{A}$ only.

Table 2: Technical key properties of biosignal amplifiers for BCI operation.

	g.BSamp	g.MOBilab+	g.USBamp
Signal type	EEG/EP/EXG	EEG/EP/EXG	EEG/EP/EXG/ECOG
Channels number N	8/16	8	16
Stackable	32-80	-	32-256
Sampling rate [Hz]	250kHz/N	256	64-38.4 k
Simultaneous sample and hold	No	No	Yes
ADC inside amplifier	No	Yes	Yes
#ADCs	1	1	16
ADC resolution [Bit]	16	16	24
Over sampling	-	-	19.400 at 128 Hz
Conversion time [μs]	4 μs	43 μs	26 μs
Time delay between 1st and last channel	Conversion time * N	Conversion time * 8	Conversion time
Interface	PCI/PCMCIA	Bluetooth	USB 2.0
Range [m]	2	30	3
Power supply	12 V medical power supply or battery	4 AA batteries	5 V medical power supply or battery
Operation time on battery [h]	8	36	8
Input Sensitivity	$\pm 5 \text{ mV}/\pm 500 \mu\text{V}$	$\pm 500 \mu\text{V}$	$\pm 250 \text{ mV}$
Minimum high pass [Hz]	0.01	0.01	0
Maximum low pass [Hz]	5 k	100	6.8 k
Band pass filter	Analog	Analog	Digital (DSP)
Notch filter	Analog	-	Digital (DSP)
Derivation	Bipolar	Monopolar	Monopolar/bipolar (DSP)
# ground potentials	2	1	4
Connectors	1.5 mm safety	1.5 mm safety / system connectors	1.5 mm safety / system connectors
Impedance test	External	External	Internal
Applied part	CF	BF	CF
Isolation of applied part	Isolation amplifier	Opto-coupler	Opto-coupler
Digital I/Os	8	8	8
Scanned with inputs	No	Yes	Yes

The next important feature is the number of electrodes used. ECG recordings are mostly done with 1 bipolar derivation or a 12 lead ECG derivation (Goldberger, Einthoven, Wilson) is performed. EMG recordings are made mostly made with 1-10 derivations. EOG is recorded with one diagonal bipolar channels around the eye or with one horizontal and one vertical channels around each eye. Additionally one amplifier channels is needed for each external sensor that should be recorded (respiration, flow, temperature, acceleration, velocity, oxygen saturation,...). For slow wave approaches or oscillations in the alpha and beta range and P300 systems, a total of 1-8 EEG channels are sufficient (Ref. 10 - Ref. 12). BCIs that use spatial filtering, such as common spatial pattern (CSP), require more channels (16-128) (Ref. 13). For ECoG recordings, 64-128 channel montages are typically used (Ref. 3). Therefore, stack-able systems might be advantageous because they can extend the functionality with future applications. A stack-able e.g. 64 channel system can also be split into four 16 channels systems if required for some experiments. USB 2.0 provides a much higher bandwidth than Bluetooth and therefore allows higher sampling frequencies and more channels. Two clear advantages of Bluetooth devices are portability and mobility. Subjects can move freely within a radius of about 30 meters. USB based wired devices have cable lengths of about 3 meters, and the distance between a stand-alone amplifier and a DAQ board should be as short as possible (<2 m). Another advantage is that moving the ADC as close as possible to the amplification unit yields a higher signal to noise ratio.

Amplifiers with bipolar inputs use typically instrumentation amplifiers as input units with a high common mode rejection ratio (CMRR). The electrodes are connected to the plus and the minus inputs of the amplifier and electrodes are mounted on the scalp in a distance of about 2.5 – 10 cm (see Figure 11A). The ground electrode is mounted e.g. on the forehead. Bipolar derivations have the advantage of suppressing noise and artifacts very well, so that only local brain activity near the electrodes is picked up. In contrast, monopolar input amplifiers have a common reference electrode that is typically mounted at the ear lobes or mastoids (Figure 11B). Monopolar recordings refer measurements to the reference electrode and are more sensitive to artifacts, but make it possible to calculate bipolar, small/large Laplacian, Hjorth's, or common average reference (CAR) derivations afterwards (Ref. 14). Typically, bipolar derivations are preferred if only a few channels should be used, while monopolar channels are used for recordings with many electrodes such as ECoG or when spatial filters are applied (Figure 12). Groups of ground potential separated amplifiers allow the simultaneous acquisition of other biosignals like ECG and EMG along with EEG without any interference. Another benefit of separated ground potentials is the ability to record multiple subjects with one amplifier, which allows e.g. BCI games where users can play against each other (Ref. 15).




Figure 11: Scalp electrode montage with bipolar (A) and monopolar recordings (B)



Figure 12: ECoG electrode with 64 electrodes

(picture from Gerwin Schalk Wadsworth Center, USA, Kai Miller and Jeff Ojemann University of Washington, USA)

The signal type (EEG, ECoG, evoked potentials – EP, EMG, EOG) also influences the necessary sampling frequency and bandwidth of the amplifier. For EEG signals, sampling frequencies of 256 Hz with a bandwidth of 0.5 – 100 Hz are typically used (Ref. 4). For ECoG recordings, sampling frequencies of 512 or 1200 Hz are applied with a bandwidth of 0.5 – 500 Hz (Ref. 3). A special case is slow waves, where a lower cut – off frequency of 0.01 Hz is needed (Ref. 10). For P300 based systems, a bandwidth of 0.1 – 30 Hz is typically used (Ref. 16). Notch filters are used to suppress the 50 Hz or 60 Hz power line interference. A notch filter is typically a narrow band-stop filter having a very high order. Digital filtering has the advantage that every filter type (Butterworth, Bessel, etc), filter order, and cut-off frequency can be realized. Analog filters inside the amplifier are predefined and can therefore not be changed. The high input range of g.USBamp of ± 250 mV combined with a 24 Bit converter (resolution of 29 nV) allows measuring all types of biosignals (EMG, ECG, EOG, EPs, EEG, ECoG) without changing the amplification factor of the device. For 16 Bit AD converters, the input range must be lower in order to have a high enough ADC resolution.

3.2 Programming Environment

Physiological recording systems are constructed under different operating systems (OS) and programming environments. Windows is currently the most widely distributed platform, but there are also implementations under Window Mobile, Linux and Mac OS. C++, LabVIEW (National Instruments Corp., Austin, TX, USA) and MATLAB (The MathWorks Inc., Natick, USA) are mostly used as programming languages. C++ implementations have the advantages that no underlying software package is needed when the software should be distributed, and allow a very flexible system design. Therefore, a C++ Application Program Interface (API) was developed that allows the integration of the amplifiers with all features into programs running under Windows or Windows Mobile. The main disadvantage is the longer development time.

Under the MATLAB environment, several specialized toolboxes such as signal processing, statistics, wavelets, and neural networks are available, which are highly useful components for a BCI system. Signal processing algorithms are needed for feature extraction, classification methods are needed to separate biosignal patterns into distinct classes, and statistical functions are needed e.g. for performing group studies. Therefore, a MATLAB API was also developed, which is seamlessly integrated into the Data Acquisition Toolbox. This allows direct control of the amplification unit from the MATLAB command window to capture the biosignal data in real-time

and to write user specific m-files for the data processing. Furthermore, standard MATLAB toolboxes can be used for processing, as well as self-written programs. The MATLAB processing engine is based upon highly optimized matrix operations, allowing very high processing speed. Such a processing speed is very difficult to realize with self-written C code.

4 Signals and signal processing algorithms

In the following this document gives a literature overview about existing and well known signal processing algorithms applied for the analysis of different biosignals measurable with the systems presented above. Due to the enormous amount of different biosignals only those are referred which are in some sense related to the tasks of BrainAble and therefore usable within the project. The following signals are being considered within this document.

- ECG
- EEG
- EMG
- Electrodermal activity
- Respiration
- Acceleration
- Video tracking for movement capturing

We would like to stress that although there are many publications regarding other famous sensors, like *e.g.* SPO2 – sensors or skin temperature sensors, they are not considered here. This is mainly due to the fact that its normal use is beyond the BrainAble scope. The main use of SPO2 – sensors for example is to analyze the effect of high altitudes or in pathologies. The same is for skin temperature which is mainly interesting for pathologies.

4.1 ECG

4.1.1 Heart Rate (HR)

The Heart Rate (HR) is usually derived from the ECG, by an algorithm detecting the single heart beats (see *e.g.* Ref. 47). This can be easily done, as each heart beat is accompanied in the ECG by a QRS complex. After detection, an optical inspection of the detected complexes is necessary to prevent false positive detections.

The inverse of the time difference between the so-called normal heart beats (QRS complexes resulting from sinus node depolarizations) gives the heart rate. The unit of the HR is defined in beats per minute (bpm). For calculating the HR, it is sufficient to use three electrodes to obtain Einthoven I or Einthoven II leads. The optimal range for the sampling frequency is above 250 Hz as a lower value would produce jitter which alters the spectrum considerably (Ref. 39). The heart rate is then sampled between consecutive intervals (NN intervals), for example as the instantaneous heart rate, IHR (Ref. 68). For deriving the IHR the value of each NN interval (in bpm) remains constant during the whole duration of its corresponding interval, hence the IHR is sampled as a step function.

Several publications describe the influence of mental tasks onto HR changes. An increase was found during cognitive processing (Ref. 36). For motor tasks the preparation leads to an HR decrease (Ref. 35, Ref. 41, Ref. 69), while it increases during mental execution (Ref. 38 and Ref. 54).

An attempt for using the HR to control a BCI was done by Scherer et al. (Ref. 61). In this paper, the HR was used to switch on and off a SSVEP controlled prosthesis. Brisk inspiration, performed by the subjects, led to an increase in the HR. Each time the HR-increase exceeded a predefined subject specific threshold, the device was switched on or off.

4.1.2 Heart Rate Variability (HRV)

Heart rate variability (HRV) describes the changes of the HR over time. HRV parameters can be divided into time domain and frequency domain measures.

4.1.2.1 Time domain methods

The time domains methods are grouped into statistical methods (SDNN, SDANN, RMSSD, NN50, pNN50) and geometric methods (HRV triangular index, TINN, differential index, logarithmic index).

Statistical method

The SDNN is the standard deviation of the NN intervals, and is calculated in many studies over for a 24h period. Comparison of SDNN, derived from measurements of different duration is not suitable. When evaluating short periods (usually 5 min.) of one measurement, the SDANN gives the standard deviation of the average NN intervals, and the SDNN index the mean of the standard deviations. Also measured are the square root of the mean squared differences of successive NN intervals (RMSSD), and the number of interval differences of successive NN intervals greater than 50 ms (NN50). The latter one depends on the length of the measured data, therefore the pNN50 is better comparable, as it is calculated by dividing the NN50 by the total number of NN intervals.

Geometric methods

The HRV triangular index and the triangular interpolation of NN (TINN) are both based on the sample density distribution D . It assigns the number of equally long NN intervals to each value of their lengths. Now, the HRV triangular index is calculated by dividing the area of D by the maximum of the distribution. The triangular interpolation of NN (TINN) is the baseline width of the minimum square difference triangular interpolation of the highest peak of the histogram of all NN intervals.

The differential index is defined as the difference between the widths of the histogram of differences between adjacent NN intervals measured at selected heights (*e.g.*, at the levels of 1000 and 10 000 samples) (Ref. 29) and the logarithmic index is the coefficient ϕ of the negative exponential curve $k \cdot e^{-\phi t}$, which is the best approximation of the histogram of absolute differences between adjacent NN intervals (Ref. 60).

4.1.2.2 Frequency domain methods

The frequency domain parameters are all derived from an estimation of the power spectral density (PSD). This estimation can be done via FFT or parametric methods like an autoregressive model. Following this calculation one can extract several power components out of this estimation.

- Ultra low frequency (ULF): 0 Hz – 0.0033 Hz
- Very low frequency (VLF): 0.0033 Hz – 0.04 Hz

- Low frequency (LF): 0.04 Hz – 0.15 Hz
- High frequency (HF): 0.15 Hz – 0.4 Hz.
- LF/HF

The ULF is only used when recording 24-hour data. The VLF is influenced by parasympathetic activity (Ref. 64), while the LF is driven by both, the sympathetic and parasympathetic system. HR oscillations within this band are most likely a baroreflex response to the 10-s blood pressure oscillations (Ref. 32, Ref. 31). The HF band obtains its influence by the respiratory sinus arrhythmia (RSA) that is a heart rate oscillation driven by respiration (Ref. 70, Ref. 71, Ref. 44). The ratio LF/HF describes the balanced behavior of the sympathetic and parasympathetic systems and is therefore an indicator to see which of the two systems is actually the dominant one. When the LF component is increased and the HF component is decreased the subject is suffering mental stress or is *e.g.* at high altitude.

Even with little subjective awareness of the reduced amount of oxygen at an altitude of 2700 m, the cardiovascular and central nervous system are already affected. A study on the Dachstein showed that the HR increased from 990m altitude to 2700m altitude in a group of 10 subjects. Additionally, heart-rate variability (HRV) parameters were decreased significantly. Furthermore, with the increase in altitude, the sympathetic system becomes more active compared to the parasympathetic system (Ref. 72).

The HR, the HRV and the event-related heart rate changes were calculated from the acquired ECG data in social interaction VR simulations (Ref. 73). The study shows that the HR and HRV parameters vary significantly between the baseline and social interaction experiments. Event-related HR changes show the occurrence of breaks in presence (VR projection switched off) and also signified the virtual character utterances.

4.2 EEG

The so called ongoing EEG including brain waves or oscillations categorized into different frequency bands can be seen as a common activity of a large population of neurons in the neocortex.

Amplitudes of the ongoing EEG are in the range of 50 – 100 μ V. The amplitudes depend on the type of EEG derivation (bipolar derivation yields smaller amplitudes compared to monopolar derivations) and on the location of electrode placement. The interesting frequency ranges are between 0 – 40 Hz. Sometimes components up to 80 Hz are investigated. Table 3 gives an overview over the used frequency bands in EEG analysis. Figure 13 displays an example of typical EEG traces in different frequency ranges.

Table 3: Frequency range of the different EEG bands

EEG band		Frequency range [Hz]
Delta		0.5 – 4
Theta		4 – 8
Alpha	Lower alpha	8 –10
	Upper alpha	10 – 13
Beta		13 –30
Gamma		> 30




Figure 13: Typical EEG traces for oscillatory components in different frequency bands. Theta activity occurs in children and sleeping adults and delta activity in infants and sleeping adults. Alpha activity is best observed in occipital regions and beta activity can be seen if alpha rhythmic activity disappears mainly in parietal and frontal areas in adults. BCI system can be realized with slow waves (Ref. 6), motor imagery (Ref. 6), SSVEP (Ref. 28, Ref. 53) or the P300 evoked potential (Ref. 16).

4.2.1 Slow Cortical Potentials

Potential shift of the scalp EEG over 0.5 – 10 s are called slow cortical potentials. Reduced cortical activation goes ahead with positive SCPs, while negative SCPs are associated with movement and other functions involving cortical activation (Ref. 74). People are able to learn how to control these potentials, hence it is possible to use them for BCIs as Birbaumer and his colleagues did (Ref. 6, Ref. 10, Ref. 88). The main disadvantage of this method is the extensive training time to learn how to control the SCPs. Users need to train in several 1-2 h sessions/week over weeks or months (Ref. 74).

4.2.2 P300

The P300 wave was first discovered by Sutton (Ref. 63). It elicits when an unlikely event occurs randomly between events with high probability. In the EEG signal the P 300 appears as a positive wave 300 ms after stimulus onset. The electrodes are placed over the posterior scalp.

Its main usage in BCIs is for spelling devices, but one can also use it for control tasks (for example games (Ref. 40) or navigation (e.g. to move a computer-mouse (Ref. 33)). When using P300 as a spelling device, a matrix of characters is shown to the subject. Now the rows and columns (or in some paradigms the single characters) of the matrix are flashing in random order, while the

person concentrates only on the character he wants to spell. For better concentrating, it is recommended to count how many times the character flashes. Every time the desired character flashes, a P300 wave occurs. As the detection of one single event would be imprecise, more than one trial (flashing of each character) has to be carried out to achieve a proper accuracy.

Krusienki et al. (Ref. 48) evaluated different classification techniques for the P300 speller, wherein the stepwise linear discriminant analysis (SWLDA) and the Fisher's linear discriminant analysis provided the best overall performance and implementation characteristics.

A recent study (Ref. 43), performed on 100 subjects, revealed an average accuracy level of 91.1%, with a spelling time of 28.8 s for one single character. Each character was selected out of a matrix of 36 characters.

4.2.3 SSVEP

Steady state visual evoked potentials (SSVEP)-based BCIs use several stationary flashing sources (e.g. flickering LEDs, or phase-reversing checkerboards), each of them flashing with another constant frequency. When a person gazes at one of these sources, the specific frequency component will increase in the measured EEG, over the occipital lobe. Hence, when using different light sources, each of them representing a predefined command, the person gives this command via gazing onto the source. The classification is either done by FFT-based spectrum comparison, preferably including also the harmonics (Ref. 53), or via the canonical correlation analysis (CCA) (see Ref. 49). A third possibility is via the minimum energy approach which was published by O. Friman et.al. in 2007 (Ref. 75) and requires no training.

Typical SSVEP applications are made for navigation, for example Middendorf et al. (Ref. 52) used SSVEPs to control the roll position of a flight simulator. The number of classes varies between two and eight, although Gao et al. (Ref. 42) established an experiment with even 48 targets, but in this experiment they had only one subject. Bakardijan et al. (Ref. 26) investigated SSVEP responses for frequencies between 5 and 84 Hz to finding the strongest response between 5.6 Hz and 15.3 Hz peaking at 12 Hz. With their frequency-optimized-eight-command BCI they achieved a mean success rate of 98 % and an information transfer rate (ITR) of 50 bits/min. Bin et al. (Ref. 28) reports of a six-target BCI with an average accuracy of 95.3% and an information transfer rate of 58 ± 9.6 bits/min.

Although most SSVEP-based BCIs work with gaze shifting towards a source, recent studies (Ref. 25, Ref. 67) proofed that only selective attention onto a pattern alone is sufficient for control. The latter paper achieved an overall classification accuracy of $72.6 \pm 16.1\%$ after 3 training days. Therefore also severely disabled people, who are not able to move their eyes, can control an SSVEP-based BCI.

4.2.4 Motor imagery

When subjects perform or only imagine motor tasks, an event related desynchronization (ERD) (Ref. 57, Ref. 58) and an event related synchronization (ERS) is detectable by changes of EEG rhythms on positions close to the respective sensorimotor areas. The ERD is indicated by a decrease of power in the upper alpha band and lower beta band, starting 2 seconds before movement onset on the contralateral hemisphere and becomes bilaterally symmetrical immediately before execution of movement (Ref. 56). An ERS appears either after termination of the movement, or simultaneously to the ERD, but in other areas of the cortex. The decrease/increase is always measured in comparison to the power in a reference interval, for

example a few seconds before the movement occurs. For classification there are several approaches used. The simplest one is by calculating the bandpower in a specific frequency band and consecutive discrimination via a Fisher linear discriminant analysis. Other classification strategies are support vector machines (SVM) (e.g. Ref. 80), principal component analysis (PCA) (Ref. 76), or common spatial patterns (CSP) (Ref. 77)

4.3 EOG and eye movements

The human eye acts as a dipole (the cornea is positive, the fundus negative). Hence it is simple to measure eye-movements by placing a pair of electrodes on the left and right side of the eye (near the external canthus) and another pair closely below and above it. Following an eye movement a DC shift will occur between the electrodes. The EOG has amplitude of about $20\mu\text{V}$ per degree of rotation. A frequency response of up to 30 Hz is adequate (Ref. 78). Beneath eye movement, also blinking causes a DC shift in the signal.

The EOG is used in behavioral studies (e.g. sleep research), or in EEG measurements to identify eye movement artifacts. The importance of measuring and dealing correctly with EOG artifacts was carried out by Fatourehchi et al. (Ref. 81) when they analyzed more than 250 papers, and revealed weaknesses of these studies, considering EOG and EMG artifacts. The ways of dealing with artifacts are either a manual or automatic rejection of the data, or automatic removal of the artifact with filtering, PCA, ICA or regression (Ref. 81).

Eye movements can also be recorded with video eye-trackers where the cameras are positioned close to the eye. The advantage is that no electrodes must be assembled onto the subject, but the eyes must be visible by the camera. Eye tracking was recently used to control the movement of a car and for spelling devices that allows selecting characters on the screen.

4.4 EMG

The EMG can be easily recorded by means of either surface electrodes, placed on the skin, or needle electrodes inserted into the muscle (Ref. 78). The frequency response has a wide range, being maximal at frequencies higher than 30 Hz (Ref. 82; Ref. 83). The EMG can be used to monitor motor activities, such as voluntary foot movement and also spasm or epileptic seizures. Difficult tasks may cause an increase in EMG activity related to the movement of facial muscles (Ref. 84, Ref. 85). Fatigue can be detected by the degradation of the facial muscle activity (Ref. 90, Ref. 91), moreover facial EMG can recognize the facial expression, such as pleasure, anger and sorrow (Ref. 91).

When measuring EEG it is important, to remove EMG related artifacts, for example McFarland et al. examined the presence and characteristics of EMG contamination during new users' initial brain-computer interface (BCI) training sessions (Ref. 83). The methods for dealing with artifacts are the same as for removing EOG artifacts (Ref. 81). EMG can also be used as an input device for human computer interaction allowing to control prosthetics, and orthotic devices.

4.5 Electrodermal activity (EDA)

There are two parameters that can be monitored with the EDA (Ref. 30, Ref. 65), the Skin conductance level (SCL) and the Skin Conductance Response (SCR).

The range of the skin resistance among subjects is from kiloohm to megaohms. Transient responses, related to sudden changes in psychological state are on the order of 100 ohms (Ref. 78).

The SCL describes the overall amount of sympathetic arousal, while the SCR reflects transient changes in conductance (Ref. 73). A sampling frequency of 32 Hz is sufficient. Because of the inter-individual variation of conductance a baseline recording needs to be done before each measurement.

Slater (Ref. 73) used GSR (Galvanic Skin Response is another wording for SCR) and HR to quantify breaks-in presence (shut off of the VR simulation during the experiment). The frequency response of the GSR signal was calculated with a wavelet analysis. GSR parameters and event-related heart rate changes show the occurrence of breaks in presence. There were also differences in response observed participants who reported more or less social anxiety.

4.6 Respiration

Persons perform respiration by either using the rib cage or by abdominal movement. Therefore, when using an elastic belt for recording, it is important to place it according to each person, or to use two belts to measure both movements. When the signal is measured with a thermal sensor, it has to be placed that way to measure the airflow from both, the nose and the mouth.

4.7 Tracking body movements of users

The measurement of acceleration and the position in space of e.g. a hand-held device can be used to control devices, to monitor limb movements, or overall activity of persons. Several techniques are used to track the body movements of the users. These could be either video-based tracking systems or the systems that use accelerometers (e.g. Wii control).

4.7.1 Video based tracking in VR applications (initiated by UPF)

It has been shown that the type of tracking technology chosen to enable interactions within virtual environments can have a great impact on the behavior and the personal experience of users (Ref. 94, Ref. 95). In particular, the impact of the use of full body interaction in virtual reality (VR) experiences is expanding area of current research and many real-time systems have been developed to enable full body interaction within virtual reality environments.

One of the pioneering systems is the ALIVE system, which used a vision-based interface to interpret the actions of users to trigger specific behaviors in semi-autonomous virtual agents (Ref. 96). This system provided more complex and new experiences as opposed to traditional VR systems. Many vision-based systems exploit full-body motion to interact with interactive environments (Ref. 97), and recently some research has been done to assess the affective content of gesture (Ref. 92; Ref. 93).

UPF-SPECS have been working and developing several custom video-based tracking systems. For example, in Rehabilitation Gaming System - RGS (specs.upf.edu/rgs) project the movements of the upper extremities are tracked using the custom vision based tracking system AnTS, developed by UPF-SPECS. AnTS is a general purpose multiple object tracking tool based on Bayesian inference that contains a number of filters and color tracking methods as well as lens distortion and perspective correction techniques that make it well suited for this task (Ref. 98). In RGS AnTS tracks unique colored patches placed at the wrists and elbows of the user. In this way the visual

segmentation task is easily resolved and potential ambiguities due to the crossing of the upper extremities are avoided.

4.7.2 Hand-held devices for gesture tracking

There are a lot of possible off the shelf hand-held devices which can be used for gesture tracking. As representatives of possible devices only three will be discussed within this state of the art report:

- Wii Remote
- Tyromotion Pablo
- F.A.B. System

4.7.2.1 *Wii Remote*

The Wii Remote (Nintendo Co., Ltd., Kyoto, Japan) uses two sensor techniques: an optical sensor to determine the pointing direction of the controller, and an accelerometer, able to distinguish between three spatial axes. The Wii console uses a sensor bar, consisting of 10 infrared LEDs, that has to be placed in the remote's pointing direction. For the use without the console, however, one can use also for example a string of infrared LEDs. The optical sensor in the Wii Remote can then detect these LEDs and calculate via triangulation for example the distance to the bar.

The detection works well within a distance of five meters, with a maximum distance of ten meters. The optical sensor is able to detect the distance to the sensor bar as well as up-down and left-right movements. With the accelerometer one can monitor fast movements and rotations. Additionally there are several buttons on the Wii Remote. There exist also several attachments such as the Nunchuk, which contains an accelerometer, an analog stick and two trigger buttons.

A study of Dale et al. (Ref. 34) successfully utilized the Wii Remote via the open source package DarwiinRemote (Ref. 37) to monitor subjects arm movements on an Apple Macintosh computer. A similar package, called WiinRemote (Ref. 66), is available for Microsoft Windows. The connection to the Wii console, or a computer is established via Bluetooth.

4.7.2.2 *Tyromotion Pablo*

With the Pablo system, developed by Tyromotion (Tyromotion GmbH, Graz, Austria) it is possible to measure both, the force of flexion/extension of the hand, as well as the movement of the upper extremities and its joints.

4.7.2.3 *F.A.B. System*

The following description is a citation from BIOSYN Systems Homepage (<http://www.biosynsystems.net/f-a-b-system>)

The FAB System (Functional Assessment of Biomechanics™) developed by BIOSYN SYSTEMS INC. is the world's first full body wireless motion capture system based on "camera free" inertial sensor technology. This innovative technology allows completely free movement without the need to stay in view of cameras.

FAB is a completely portable system and can be packed up and transported in a standard size briefcase. FAB only requires a data logger or laptop, and up to 17 small sensors which are

attached to user selectable body segments. These sensors can be attached under or above clothing.

Data can be recorded in real time with direct connection to your laptop or recorded using a handheld Data Logger for post recording data analysis and animation. This makes FAB the most portable motion capture system available.

FAB is used in many industries including sports medicine, clinical and industrial applications. It has been used for athletic training, physiotherapy, occupational therapy, ergonomic studies, industrial design, and the entertainment industry.



5 References

- Ref. 1 G. Pfurtscheller, C. Neuper, D. Flotzinger, and M. Pregenzer, EEG-based discrimination between imagination of right and left hand movement. *Electroenceph. clin. Neurophysiol.*, 1997. 103: p. 642-651.
- Ref. 2 C. Guger, H. Ramoser, and G. Pfurtscheller, Real-time EEG analysis with subject-specific spatial patterns for a Brain Computer Interface (BCI). *IEEE Trans Neural Syst Rehabil Eng.*, 2000. 8: p. 447-456.
- Ref. 3 E.C. Leuthardt, G. Schalk, J.R. Wolpaw, J.G. Ojemann, and D.W. Moran, A brain-computer interface using electrocorticographic signals in humans. *J. Neural Eng.*, 2004. 1: p. 63-71.
- Ref. 4 C. Guger, A. Schlögl, C. Neuper, D. Walterspacher, T. Strein, and G. Pfurtscheller, Rapid prototyping of an EEG-based brain-computer interface (BCI). *IEEE Trans. Rehab. Engng.*, 2001. vol. 9 (1): p. 49-58.
- Ref. 5 G.R. Muller-Putz, R. Scherer, C. Brauneis, and G. Pfurtscheller, Steady-state visual evoked potential (SSVEP)-based communication: impact of harmonic frequency components. *J. Neural Eng*, 2005. 2(4): p. 123-130.
- Ref. 6 N. Birbaumer, N. Ghanayim, T. Hinterberger, I. Iversen, B. Kotchoubey, A. Kubler, J. Perelmouter, E. Taub, and H. Flor, A spelling device for the paralysed. *Nature*, 1999. 398(6725): p. 297-298.
- Ref. 7 M. Thulasidas, G. Cuntai, and W. Jiankang, Robust classification of EEG signal for brain-computer interface. *IEEE Trans Neural Syst Rehabil Eng.*, 2006. 14(1): p. 24-29.
- Ref. 8 G. Pfurtscheller, C. Neuper, C. Guger, B. Obermaier, M. Pregenzer, H. Ramoser, and A. Schlögl, Current trends in Graz brain-computer interface (BCI) research. *IEEE Trans. Rehab. Engng.*, 2000. vol. 8: p. 216-219.
- Ref. 9 C. Guger, Real-time data processing under Windows for an EEG-based brain-computer interface. Dissertation, University of Technology Graz, 1999.
- Ref. 10 N. Birbaumer, A. Kubler, N. Ghanayim, T. Hinterberger, J. Perelmouter, J. Kaiser, I. Iversen, B. Kotchoubey, N. Neumann, and H. Flor, The thought translation device (TTD) for completely paralyzed patients. *IEEE Trans.Rehabil.Eng*, 2000. 8(2): p. 190-193.
- Ref. 11 D.J. Krusienski, E.W. Sellers, F. Cabestaing, S. Bayouth, D.J. McFarland, T.M. Vaughan, and J.R. Wolpaw, A comparison of classification techniques for the P300 Speller. *J. Neural Eng*, 2006. 3(4): p. 299-305.
- Ref. 12 C. Guger, G. Edlinger, W. Harkam, I. Niedermayer, and G. Pfurtscheller, How many people are able to operate an EEG-based brain computer interface? *IEEE Trans. Rehab. Engng.*, 2003. vol. 11: p. 145-147.
- Ref. 13 H. Ramoser, J. Muller-Gerking, and G. Pfurtscheller, Optimal spatial filtering of single trial EEG during imagined hand movement. *IEEE Trans Neural Syst Rehabil Eng.*, 2000. 8(4): p. 441-446.
- Ref. 14 D.J. McFarland, W.A. Sarnacki, and J.R. Wolpaw, Brain-computer interface (BCI) operation: optimizing information transfer rates. *Biol.Psychol.*, 2003. 63(3): p. 237-251.


- Ref. 15 G. Edlinger, G. Krausz, F. Laundl, I. Niedermayer, C. Guger, Architectures of laboratory-PC and mobile pocket PC brain-computer interfaces. 2nd International IEEE EMBS Conference on Neural Engng, 2005. p. 120 - 123.
- Ref. 16 E.W. Sellers, D.J. Krusienski, D.J. McFarland, T.M. Vaughan, and J.R. Wolpaw, A P300 event-related potential brain-computer interface (BCI): the effects of matrix size and inter stimulus interval on performance. *Biol.Psychol.*, 2006. 73(3): p. 242-252.
- Ref. 17 G. Klem, H. Lüders, H. Jasper, and C. Elger, The ten-twenty electrode system of the International Federation. *The International Federation of Clinical Neurophysiology*. Cleveland Clinic Foundation, 1999. 52: p. 3-6.
- Ref. 18 B. Obermaier, C. Guger, C. Neuper, and G. Pfurtscheller, Hidden Markov Models for online classification of single trial EEG data. *Pattern recognition letters* 22, 2001: p. 1299-1309.
- Ref. 19 C.Neuper, G. Pfurtscheller, C. Guger, B. Obermaier, M. Pregenzer, H. Ramoser, and A. Schlögl, Current trends in Graz brain-computer interface (BCI) research. *IEEE Trans. Rehab. Engng.*, 2000. vol. 8: p. 216-219.
- Ref. 20 G. Cuntai, M. Thulasidas, and W. Jiankang. High performance P300 speller for brain-computer interface.
- Ref. 21 M. Waldhauser, Offline and online processing of evoked potentials. Master thesis, FH Linz, 2006.
- Ref. 22 E.W. Sellers and E. Donchin, A P300-based brain-computer interface: initial tests by ALS patients. *Clin.Neurophysiol.*, 2006. 117(3): p. 538-548.
- Ref. 23 C. Guger, C. Groenegress, C. Holzner, G. Edlinger, and M. Slater. Brain-computer interface for controlling Virtual Environments. in 2nd international conference on applied human factors and ergonomics. 2008. Las Vegas, USA.
- Ref. 24 Akselrod. (1981). Akselrod S, Gordon D, Ubel FA, Shannon DC, Barger AC, Cohen RJ. *Science* , 220--222.
- Ref. 25 Allison, B. Z., McFarland, J., D., Schalk, G., Zheng, S. D., M., J. M., et al. (2008). Towards an independent brain-computer interface using steady state visual evoked potentials. *Clin Neurophysiol* , 399-408.
- Ref. 26 Bakardjian, H., Tanaka, T., & Cichocki, A. (2010). Optimization of SSVEP brain responses with application to eight-command Brain-Computer Interface. *Neurosci. Lett.* , 34-8.
- Ref. 27 Bigger JT, F. J. (1992). Frequency domain measures of heart period variability and mortality after myocardial infarction. *Circulation* , 85:164.
- Ref. 28 Bin, G., Gao, X., Yan, Z., Hong, B., & Gao, S. (2009). An online multi-channel SSVEP-based brain-computer interface using a canonical correlation analysis method. *Journal of Neural Engineering* , 6pp.
- Ref. 29 Bjökander I, H. C. (1992). Heart rate variability in patients with stable angina pectoris. *Eur Heart J.* , 379.
- Ref. 30 Boucsein, W. (1992). *Electrodermal Activity*. New York: Plenum Press.
- Ref. 31 C., J. (2006). The enigma of Mayer waves: facts and models. *Cardiovasc. Res.* , 12-21.
- Ref. 32 Cevese A., G. G. (2001). Baroreflex and oscillation of heart period at 0.1 Hz studied by alpha-blockade and cross-spectral analysis in healthy humans. *J. Physiol* , 235–244.
- Ref. 33 Citi, L., Poli, R., Cinel, C., & Sepulveda, F. (2008). P300-based BCI mouse with genetically-optimized analogue control. *IEEE Transactions on Neural Systems and Rehabilitation Engineering* , 51-61.
- Ref. 34 Dale R., R. J. (2008). Exploring Action Dynamics as an Index of Paired-Associate Learning. *PLoS ONE* 3(3) , e1728.

- Ref. 35 Damen, E., & Brunia, C. (1987). Changes in heart rate and slow brain potentials related to motor preparation and stimulus anticipation in a time estimation task. *Psychophysiology* , 700-713.
- Ref. 36 Danilova, N., Korshunova, S., & Sokolov, E. (1994). Indexes of heart-rate during solving arithmetical tasks in humans. *Zh. Vyssh. Nerv. Deyat* , 932-943.
- Ref. 37 DarwiinRemote. (n.d.). Retrieved from <http://sourceforge.net/projects/darwiin-remote/>
- Ref. 38 Decety, J., Jeannerod, M., Germain, M., & Pastene, J. (1991). Vegetative response during imagined movement is proportional to mental effort. *Behav. Brain Res.* , 1-5.
- Ref. 39 Electrophysiology, T. F. (1996). Heart rate variability: standards of measurement, physiological interpretation and clinical use. Task Force of the European Society of Cardiology and the North American Society of Pacing and Electrophysiology. *Circulation* , 1043--1065.
- Ref. 40 Finkea, A., Lenhardt, A., & Ritter, H. (2009). The MindGame: a P300-based brain-computer interface game. *Neural Networks* , 1329-1333.
- Ref. 41 Florian, G., Stancak, A., & Pfurtscheller, G. (1998). Cardiac response induced by voluntary self-paced finger movement. *International Journal of Psychophysiology* , 209-222.
- Ref. 42 Gao, X., Xu, D., Cheng, M., & S., G. (2003). A BCI-Based Environmental Controller for the Motion-Disabled. *IEEE Trans. Neural Sys. and Rehab. Eng.* , 137-140.
- Ref. 43 Guger, C., Daban, S., Sellers, E., Holzner, C., Krausz, G., Carabalona, R., et al. (2009). How many people are able to control a P300-based brain-computer interface (BCI)? *Neuroscience Letters* , 94-98.
- Ref. 44 Hirsh JA, B. B. (1981). Respiratory sinus arrhythmia in humans: how breathing pattern modulates heart rate. *Am J Physiol* , H620-H629.
- Ref. 45 Hon EH, L. S. (1965). Electronic evaluations of the fetal heart rate patterns preceding fetal death: further observations. *Am J Obstet Gynecol* , 814-826.
- Ref. 46 Kleiger RE, M. J.-I. (1987). Decreased heart rate variability and its association with increased mortality after acute myocardial infarction. *Am J Cardiol* , 256-262.
- Ref. 47 Köhler BU, H. C. (2002). The principles of software QRS detection. *IEEE Eng Med Biol Mag.* , 42-57.
- Ref. 48 Krusienski, D., Sellers, E., Cabestaing, F., Bayouth, S., McFarland, D., Vaughan, T., et al. (2006). A comparison of classification techniques for the P300 Speller. *Journal of Neural Engineering* , 299-305.
- Ref. 49 Lin Z., Zhang C., Wu W., & X., G. (2006). Frequency recognition based on canonical correlation analysis for SSVEP-based BCIs. *IEEE Trans Biomed Eng.* , 2610-2614.
- Ref. 50 Luczak H, L. W. (1973). An analysis of heart rate variability. *Ergonomics* , 85-97.
- Ref. 51 Malik M, F. T. (1989). Camm AJ. Heart rate variability in relation to prognosis after myocardial infarction: selection of optimal processing techniques. *Eur Heart J.* , 1060-1074.
- Ref. 52 Middendorf, M., McMillan, G., Calhoun, G., & Jones, K. S. (2000). Brain-computer interfaces based on the steady-state visual-evoked response. *IEEE Transactions on Rehabilitation Engineering* , 211-214.
- Ref. 53 Müller-Putz, G. R., Scherer, R., Brauneis, C., & Pfurtscheller, G. (2005). Steady-state visual evoked potential (SSVEP)-based communication: impact of harmonic frequency components. *Journal of Neural Engineering* , 1-8.

- Ref. 54 Oishi, K., Kasai, T., & Maeshima, T. (2000). Autonomic response specificity during motor imagery. *J. Physiol. Anthropol. Appl. Hum. Sci.* , 255-261.
- Ref. 55 Penaz J, R. J. (1968). *Spectral Analysis of Some Spontaneous Rhythms in the Circulation*. Leipzig: Biokybernetik, Karl Marx University.
- Ref. 56 Pfurtscheller G., L. d. (1999). Event-related EEG/MEG synchronization and desynchronization: basic principles. *Clinical Neurophysiology* , 1842-1857.
- Ref. 57 Pfurtscheller, G., & Aranibar, A. (1977). Event-related cortical desynchronization detected by power measurements of scalp EEG. *Clinical Neurophysiology* , 817-826.
- Ref. 58 Pfurtscheller, G., & Neuper, C. (1997). Motor imagery activates primary sensorimotor area in humans. *Neuroscience Letters* , 65-68.
- Ref. 59 Sayers, B. (1973). Analysis of heart rate variability. *Ergonomics* , 17-32.
- Ref. 60 Scherer P, O. J.-W. (1993). Definition of a new beat-to-beat parameter of heart rate variability. *PACE Pacing Clin Electrophysiol* , 16:939.
- Ref. 61 Scherer, R., Müller-Putz, G. R., & Pfurtscheller, G. (2007). Self-initiation of EEG-based brain-computer communication using the heart rate response. *Journal of Neural Engineering* , L23-L29.
- Ref. 62 Slater, M., Antley, A., Davison, A., Swapp, D., Guger, C., Barker, C., et al. (2006). A Virtual Reprise of the Stanley Milgram Obedience Experiments. *PLoS ONE* , e39.
- Ref. 63 Sutton, S., Braren, M., Zubin, J., & John, E. R. (1965). Evoked-Potential Correlates of Stimulus Uncertainty. *Science* , 1187-1188.
- Ref. 64 Taylor J.A., C. D. (1998). Mechanisms underlying very-low-frequency RR-interval oscillations in humans. *Circulation* , 547-555.
- Ref. 65 Venables, P. (1978). Psychophysiology and Psychometrics. *Psychophysiology* , 302-315.
- Ref. 66 WiinRemote. (n.d.). Retrieved from <http://onakasuita.org/wii/index-e.html>
- Ref. 67 Zhang, D., Maye, A., Gao, X., B., H., Engel, A. K., & Gao, S. (2010). An independent brain-computer interface using covert non-spatial visual selective attention. *J Neural Eng.* , 16010. Epub 2010 Jan 19.
- Ref. 68 de Boer RW, Karemaker JM, Strackee J 1985 Description of heart rate variability data in accordance with a physiological model for the genesis of heartbeats. *Psychophysiology* 22:147-155
- Ref. 69 Pfurtscheller, G., Leeb, R. and Slater, M. (2006) Cardiac responses induced during thought-based control of a virtual environment, *International Journal of Psychophysiology* 62(1), 134-140
- Ref. 70 Einbrodt, Über den Einfluß der Atembewegung auf Herzschlag und Blutdruck, *Sber. Akad. Wiss. Wien; Math. Nat. Kl., 2. Abt.*, 40 (1860), S. 361-418
- Ref. 71 Eckberg DL. Human sinus arrhythmia as an index of vagal cardiac outflow. *J Appl Physiol.* 1983;54:961-966.
- Ref. 72 Guger C., Krausert S., Domej W., Edlinger G., Tannheimer M., "EEG, ECG and oxygen concentration changes from sea level to a simulated altitude of 4000m and back to sea level" *Neuroscience Letters*, 2008 , Volume 442, Issue 2: pp 123-127
- Ref. 73 Slater M., Guger C., Edlinger G., Leeb R., Pfurtscheller G., Antley A., Garau M., Brogni A., Friedman D., "Analysis of Physiological Responses to a Social Situation in an Immersive Virtual Environment", *Presence: Teleoperators & Virtual Environments* October 2006, Vol. 15, No. 5: 553-569.
- Ref. 74 Wolpaw J. R., Birbaumer N., McFarlanda D.J., Pfurtscheller G., Vaughan T. M., "Brain-computer interfaces for communication and control", *Clinical Neurophysiology*, Volume 113, Issue 6, Pages 767-791 (June 2002)

- Ref. 75 Friman, O., Volosyak, I., Graser, A.: Multiple channel detection of Steady-State Visual Evoked Potentials for brain-Computer interfaces. *IEEE Transactions on Biomedical Engineering*, 54, 742--750, (2007).
- Ref. 76 Vallabhaneni A., "Motor imagery task classification for brain computer interface applications using spatiotemporal principle component analysis". *Neurol Res.* 2004 Apr ;26 (3):282-7 15142321 Cit:5
- Ref. 77 Müller-Gerking, J., Pfurtscheller, G., Flyvbjerg, H. (1999). „Designing optimal spatial filtering of single trial EEG classification in a movement task". *Clinical Neurophysiology*, 110:787–798, 1999.
- Ref. 78 Kamp, A., Pfurtscheller, G. Polygraphy. In: Niedermeyer, E., Lopes da Silva, F. (eds.). *Electroencephalography Basic Principles, Clinical Applications, and Related Fields*. Third Edition, 1993, Williams and Wilkins, Baltimore, Maryland.
- Ref. 79 Boucsein, W. *Electrodermal Activity*. New York: Plenum Press, 1992
- Ref. 80 Solis-Escalante T, Müller-Putz G, Pfurtscheller G. Overt foot movement detection in one single Laplacian EEG derivation. *J. Neurosci Methods* 2008 175(1):148-53
- Ref. 81 Fatourehchi M, Bashashati A, Ward RK, Birch GE. EMG and EOG artifacts in brain computer interface systems: A survey. *Clin Neurophysiol.* 2007 Mar;118(3):480-94
- Ref. 82 Anderer P, Roberts S, Schlogl A, Gruber G, Klosch G, Herrmann W, et al. Artifact processing in computerized analysis of sleep EEG – a review. *Neuropsychobiology* 1999;40:150–7.
- Ref. 83 McFarland DJ, McCane LM, David SV, Wolpaw JR. Spatial filter selection for EEG-based communication. *Electroencephalogr Clin Neurophysiol* 1997;103:386–94.
- Ref. 84 Cohen BH, Davidson RJ, Senulis JA, Saron CD, Weisman DR. Muscle tension patterns during auditory attention. *Biol Psychol*, 1992;33:133–56.
- Ref. 85 Waterink W, van Boxtel A. Facial and jaw-elevator EMG activity in relation to changes in performance level during a sustained information processing task. *Biol Psychol* 1994;37:183–98.
- Ref. 86 Venables, P. "Psychophysiology and Psychometrics." *Psychophysiology*, 1978: 302-315.
- Ref. 87 McFarland DJ, Sarnacki WA, Vaughan TM, Wolpaw JR. Brain–computer interface (BCI) operation: signal and noise during early training sessions. *Clin Neurophysiol* 2005;116:56–62.
- Ref. 88 Elbert T, Rockstroh B, Lutzenberger W, Birbaumer N. Biofeedback of slow cortical potentials. *Electroenceph clin Neurophysiol* 1980;48:293–301.
- Ref. 89 Pfurtscheller, G. & Neuper, C. Motor imagery and direct brain-computer communication *Proceedings of the IEEE*, 2001, 89, 1123-1134
- Ref. 90 Veldhuizen, I. J. T., Gaillard, A. W. K., and de Vries, J., "The influence of mental fatigue on facial EMG activity during a simulated workday," *Biological Psychology*, 63(1): 59-78, 2003.
- Ref. 91 Huang, C.N., Chen, C. H., Chung, H. Y., The Review of Applications and Measurements in Facial Electromyography, *Journal of Medical and Biological Engineering*, 25(1): 15-20, 2004.
- Ref. 92 Camurri, A., Mazzarino, B., Ricchetti, M., Timmers, R., & Volpe, G. (2004). Multimodal analysis of expressive gesture in music and dance performances. *Lecture Notes in Computer Science*, 20–39.
- Ref. 93 Camurri, A., Volpe, G., De Poli, G., & Leman, M. (2005). Communicating expressiveness and affect in multimodal interactive systems. *IEEE Multimedia*, 12(1), 43–53.

- Ref. 94 Benford, S., Greenhalgh, C., Reynard, G., Brown, C., & Koleva, B. (1998). Understanding and constructing shared spaces with mixed-reality boundaries. *ACM Transactions on Computer-Human Interaction (TOCHI)*, 5(3), 185–223.
- Ref. 95 Tang, A., Biocca, F., & Lim, L. (2004). Comparing differences in presence during social interaction in augmented reality versus virtual reality environments: An exploratory study. Paper presented at the 7th Annual International Workshop on Presence, Valencia, Spain.
- Ref. 96 Maes, P., Darrell, T., Blumberg, B., & Pentland, A. (1997). The ALIVE system: Wireless, full-body interaction with autonomous agents. *ACM Multimedia Systems*, 5(2), 105–112.
- Ref. 97 Modler, P., Myatt, T., & Saup, M. (2003). An experimental set of hand gestures for expressive control of musical parameters in real time. Paper presented at the 2003 Conference on New Interfaces for Musical Expression, Montreal, Quebec, Canada.
- Ref. 98 Mathews Z, Bermúdez i Badia S, Verschure PFMJ. A Novel Brain-Based Approach for Multi-Modal Multi-Target Tracking in a Mixed Reality Space. *INTUITION - International Conference and Workshop on Virtual Reality 2007*. Athens, Greece 2007.



6 List of Key Words/Abbreviations

EEG	Electroenzephalogram
ECoG	Electrocortigogram
ECG	Electrocardiogram
EMG	Electromyogram
EOG	Electrooculogram
EP	Evoked Potential
SSVEP	Steady State Visual Evoked Potentials
BNCI	Brain/Neuronal Computer Interface
BCI	Brain Computer Interface
HCI	Human Computer Interface
DAQ	Data Aquisition
ADC	Analog Digital Converter
DSP	Digital Signal Processing
HR	Heart Rate
HRV	Heart Rate Variability
QRS	QRS complex; part of the ECG wave form
RR	R – R intervals
NN	Normal to Normal Intervals
SDNN	Standard Deviation of NN Intervals
TINN	Triangular Index of NN Intervals
RMSSD	Root Mean Square of Squared Differences
PSD	Power Spectral Density
ULF	Ultra Low Frequency
VLF	Very Low Frequency
LF	Low Frequency
HF	High Frequency
VR	Virtual Reality
CCA	Canonical Correlation Analysis
ITR	Information Transfer Rate
ERD	Event Related Desynchronization
ERS	Event Related Synchronization
PCA	Principal Component Analysis
ICA	Independent Component Analysis
EDA	Electrodermal Activity
SCL	Skin Conductance Level
SCR	Skin Conductance Response
GSR	Galvanic Skin Response